

ZHL - Tender Document - 001 - Fabrication dated November 23, 2016

CORRIGENDUM

S. No	Item No	Item description	Remarks
1	Sec 4, General(Commercial), Conditions of contract, point no 4, delivery	4 (b): All technical assistance for installation , commissioning and monitoring of equipment shall be provided by the suppliers at no cost	Overall responsibility of fitting and ensuring that all equipments are in full working conditions belongs to the agency.
2	Technical specifications : EMT's seat	A recliner type seat to be mounted above the oxygen cylinder housing on partition wall at 370 mm from RHS wall, 430 mm maximum from floor ; EMT seat should have slogging mechanism same like normal driver seat so as to adjust according to requirement	EMT seat must be fixed.
3	7: B Type Oxygen cylinder		It is a portable cylinder
4	Sec 3: Technical bid : Cover A	EMD os Rs 5 Lakhs in form of dd/ bankers cheque/ BG to be submitted	NSIC Certificate will be accepted as EMD
5	Technical specifications , Part A Scope of fabrication work . Wall panelling work pilot and patient compartments	Since no standard for fire retardent have been specified hene we rewuest you to specify the current standard being used in railways and other transport solutions in India and hence it should be ammended as : " Wall panel work / squad bench etc . All interiors , internal slides, cover to close the wheel bump and roof should be panelled in Non Haleogenated FRP of Grade C - 1196HF as per NCD 1409 fire retardent in gel coat whote colour , finished (No Painting will be allowed)	Wall panel work / squad bench etc . All interiors , internal slides, cover to close the wheel bump and roof should be panelled in Non Haleogenated FRP of Grade C - 1196HF as per NCD 1409 fire retardent in gel coat whote colour , finished (No Painting will be allowed)
6	Performance security deposit	The succesful bidder, within 15 days of receipt of purchase order shall be required to submit security deposit of	The performance security percentage is to be 10 % of order value initially for

		10 % of order value in form of BG in favor of service provider. The performance security shall be returned to the suppliers on successful completion of contract after the expiry of the period covered under standard warranty and submission of performance for an amount equal to 5% of order value	period of STANDARD WARRANTY. Post expiry of STANDARD WARRANTY period and start of AMC period, the performance security is to be 5 % of the contract value. AMC Performance Guarantee to be submitted in BG form to the service provider, ZHL
7	Warranty period	Warranty period of 3 years and AMC/ CMC on fabrication and medical equipments is showing as 4 as well as 5 years . Pls confirm the actual AMC period after warranty	The supplier shall quote the amount for comprehensive annual maintenance of the equipment and fabrication supplied separately , after expiry of standard warranty, for a further period so as to cover the equipment and fabrication for a total of 5 years , in the price bid.
8		The fabrication is of BLS ambulance with no medical equipments so should we quote for CMC for fabrication and equipment also	The tender is for fabrication and medical equipment supply and fitment in 38 BLS ambulances
9	Technical evaluation	Prototype: 10 marks	Stands deleted from marking at technical evaluation stage.

10			The words marks and points have been used interchangeably in the tender document. It means the same
11	Clause no 2, Commercial bid	The landed price per unit including standard warranty and AMC and all taxes (MP VAT will be excluded) and duties shall be the criterion for determining the L1 rate	MPVAT will be included in the price bid. The price bid will be inclusive of all taxes
12	Technical evaluation	Prototype : 10 marks	Since prototype stands deleted from technical evaluation stage, so the 10 marks allotted for prototype is distributed evenly between experience criterion and turnover criterion (5 marks each)

13		Technical evaluation	To clarify marks will be calculated as follows : Suppose H1 turnover is 5 Crores and H2 turnover is 4 Crores , so marking for H1 will be full 25 marks and marking for H2 will be 4 crores / 5 crores * 25 (max marks for that category)= 20
14		Already available fitments and fixtures in the ambulances	The fabricator will try to use as many of the fixtures and equipments that are already available on the chassis as possible. The fabricator will sell the unused equipment and deduct it from their bid price. The bidder will quote the net price only.
15		Financial bid opening date	Due to Govt holiday in MP, financial bid opening to be held on 14 th of December instead of 13 th of December . technical bid submission and opening remains same as in original RFP, 8 th of December.
16		Protoype inspection	Prototype inspection in 15 days on site from date of award of L1 letter. The delivery of the vehicles needs to be completed within 30 days of the prototype clearance.

By order and approval of Procurement Committee for procurement of assets for integrated emergency services in MP