

**Ziqitza Health Care Limited**  
**7th Floor, IDCO Tower, Near Rupali Square, Sahid Nagar**  
**Bhubaneswar, Khurda, Odisha, PIN-751007**

**CORRIGENDUM-1**

**THE REQUEST FOR PROPOSAL (RFP) FOR SUPPLY AND INSTALLATION OF HAND-HELD PULSE OXIMETER FOR EMAS**  
**AMBULANCES**

With reference to the Tender Document no. **001, 27<sup>th</sup> April 2023**, published in the **“The Samaj” & “The Sambad”** on **28<sup>th</sup> April 2023**, and published in the **“The Times of India”** on **29<sup>th</sup> April 2023**, the last date for submission of complete bid documents is hereby extended to 17:00 PM of 25.05.2023. The technical bid and Financial bid will be opened on 26.05.2023 at 11:00 AM.

The response to the queries raised in the pre-bid meeting is as appended.

**Ziqitza Health Care Ltd.**

Bidders Name	SI No.	RFP Page No.	Tender Specification	Request for Amendment	Clarification
Rabindra Surgical Pvt. Ltd.	1	17	10. Should have Respiratory Rate Display	We do not have respiratory rate display but we have plethysmograph, Please amend it for plethysmography	No Change
	2		11. Rechargeable Battery with Charging Station/adapter backup minimum 2hours	Our system is functional with AA Disposable battery only as this is a spot check device and in ambulance or mobile usages recharging docking station is never safe. Please amendment it.	No Change
	3		Spo2 probe(Re-usable):Adult-1no	Please define the adult probe ( Sensor) and request you to amend for sensor from the OEM only for absolutely correct result. Please amendment it.	No Change
	4		Spo2 probe(Re-Usable):Paediatric-1no	Please define the paediatric probe ( Sensor) and request you to amend for sensor from the OEM only for absolutely correct result. Please amendment it.	No Change